

Certificate of Compliance

Certificate Number **100107 - E257746A**
Report Reference **E257746, December 9th, 2006**
Issue Date **2007 January 10**

Page 1 of 1



Issued to: **SOCOMEK S A**
1 RTE DE WESTHOUSE
BOITE POSTALE 10 67235 BENFELD CEDEX FRANCE FRANCE

This is to certify that representative samples of

Measuring system


model DIRIS A40 (48250A40). Optional: Module energy (48250090)
Module energy + harmonique (48250091). Module RS485. (48250092) Module 4/20
mA (48250093). Module 2IN/2OUT (48250094).
Module Profibus (48250096). Module Memory (48250097).

Have been investigated by Underwriters Laboratories Inc.® (UL) or any authorized licensee of UL in accordance with the Standard(s) indicated on this Certificate.

Standard(s) for Safety: **UL 61010-1- Electrical Equipment for Laboratory Use; Part 1: General Requirements**
CSA-C22.2 No. 61010-1 - Electrical Equipment for Laboratory Use, Part 1: General Requirements

Additional Information: **ELECTRICAL RATING:**
Voltage: 110 - 240 Vac; 50/60 Hz or 120 – 250 V dc.
Power: 10 VA.
Measuring circuit: category III for measuring line to neutral up to 300 V and category II for 300 V to 600 V; 6 A maximum.

Only those products bearing the UL Listing Mark for the US and Canada should be considered as being covered by UL's Listing and Follow-Up Service meeting the appropriate requirements for US and Canada.

The UL Listing Mark for the US and Canada generally includes: the UL in a circle symbol with "C" and "US" identifiers:  the word "LISTED"; a control number (may be alphanumeric) assigned by UL; and the product category name (product identifier) as indicated in the appropriate UL Directory.

Look for the UL Listing Mark on the product

Issued by: *Dominique Poupeau*
Dominique Poupeau, Senior Project Engineer

UL International France SA

Any information and documentation involving UL Mark services are provided on behalf of Underwriters Laboratories Inc. (UL) or any authorized licensee of UL.
For questions in France, you may call +33 (0)1 60 19 88 00.

Reviewed by: *Manfred Mueller*
Manfred Mueller, Senior Project Engineer

UL International France SA